

Ultimate Analysis CUPS.COM

Multi-Drug Urine Test Cup

Catalogue No.: UA-CDOA-L6125/UA-CDOA-L2135

CLIA CATEGORIZATION: WAIVED URINE SCREENING TEST RESULTS AT 5 MINUTES

UA Multi-Drug Urine Test Cup are competitive binding, lateral flow immunochromatographic assays for qualitative and simultaneous detection of Amphetamine, Secobarbital, Buprenorphine, Oxazepam, Cocaine, Methyleneiodoxymethamphetamine, Methamphetamine, Morphine, Methadone, Oxycodone, Phencyclidine, Nortriptyline and Cannabinoids in human urine with below cutoff concentrations and approximate detection time:

Drug (Identifier)	Calibrator	Cut-off Level	Minimum Detection Time	Maximum Detection Time
Amphetamine (AMP500)	d-Amphetamine	500 ng/mL	2-7 hours	1-2 days
Secobarbital (BAR300)	Secobarbital	300 ng/mL	2-4 hours	1-4 days
Buprenorphine (BUP10)	Buprenorphine	10 ng/mL	4 hours	1-3 days
Oxazepam (BZO300)	Oxazepam	300 ng/mL	2-7 hours	1-2 days
Cocaine (COC150)	Benzoyllecgonine	150 ng/mL	1-4 hours	2-4 days
Methyleneiodoxymethamphetamine (MDMA500)	3,4-Methyleneiodoxymethamphetamine	500 ng/mL	2-7 hours	2-4 days
Methamphetamine (MET500/mAMP500)	D(+)-Methamphetamine	500 ng/mL	2-7 hours	2-4 days
Morphine (MOP/OPI300)	Morphine	300 ng/mL	2 hours	2-3 days
Methadone (MTD300)	Methadone	300 ng/mL	3-8 hours	1-3 days
Oxycodone (OXY100)	Oxycodone	100 ng/mL	4 hours	1-3 days
Phencyclidine (PCP25)	Phencyclidine	25 ng/mL	4-6 hours	7-14 days
Nortriptyline (TCA1000)	Nortriptyline	1000 ng/mL	8-12 hours	2-7 days
Cannabinoids (THC50)	11-nor-Δ9-THC-9-COOH	50 ng/mL	2 hours	Up to 5+ days

UA Multi-Drug Urine Test Cup is intended for over-the-counter and for prescription use. For *in vitro* diagnostic use.

The tests may yield positive results for the prescription drugs Buprenorphine, Nortriptyline, Oxazepam, Secobarbital and Oxycodone when taken at or above prescribed doses. It is not intended to distinguish between prescription use or abuse of these drugs. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly in evaluating a preliminary positive result.

The tests provide only preliminary results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

WARNINGS AND PRECAUTIONS

1. The test kit is for external use only.
2. Discard after first use. The test kit cannot be used more than once.
3. Do not use the test kit beyond expiration date.
4. Do not use the test kit if the pouch is punctured or not well sealed.
5. Keep out of the reach of children.

CONTENT OF THE KIT

1. 25 test devices, each in one pouch with two desiccants. The desiccants are for storage purposes only and are not used in the test procedure.
2. One (1) Package Insert
3. 25 Security Seals

4. 25 Pieces of Gloves

MATERIAL REQUIRED BUT NOT PROVIDED

Timer or Clock

STORAGE AND STABILITY

Store at 39°F-86°F (4°C-30°C) in the sealed pouch up to the expiration date. Keep away from direct sunlight, moisture and heat. DO NOT FREEZE.

SPECIMEN COLLECTION AND PREPARATION

WHEN TO COLLECT URINE FOR THE TEST?

Collect urine specimen after minimum detection time following suspected drug use. Urine collection time is very important in detecting any drugs of abuse. Each drug is cleared by the body and is detected in the urine at different times and rates. Please refer to the minimum or maximum detection time of each drug in this instruction.

HOW TO COLLECT URINE?

1. Remove the test cup from the foil pouch by tearing at the notch. Use it as soon as possible. Instruct the donor to remove the test cup lid and void directly into the test cup until reach the Minimum Urine Level mark (approximately 25 mL). It is acceptable to collect extra volume of urine. If insufficient specimen has been collected, instruct the donor to provide urine specimen again with another new test cup. Wipe off any splashes or spills that may be on the outside of the cup. It is recommended to wear gloves when handling the test cup with urine specimen.
2. Observe the temperature strip affixed on the test cup between 2 to 4 minutes after urine is voided into the cup. The temperature between 90°F-100°F (32°C to 38°C) indicates the fresh uncontaminated sample. If the temperature is out of this range, instruct the donor to provide urine specimen again with another new test cup.

HOW TO DO THE TEST?

Test should be performed at room temperature 65°F-86°F (18°C-30°C).

1. After the urine has been collected properly, tighten the lid and place the test cup on a flat surface.
2. Peel off the label from right to left.
3. For the drug tests, read the drug test results at 5 minutes. The results can be stable for 30 minutes.



Note: Results after more than 30 minutes may be not accurate and should not be read.

READING THE RESULTS

Negative (-)

A colored band is visible in each Control Region (C) and the appropriate Test Region (T). It indicates that the concentration of the corresponding drug of that specific test zone is zero or below the detection limit of the test.


Preliminary Positive (+)

A colored band is visible in each Control Region (C). No colored band appears in the appropriate Test Region

- (T). It indicates a preliminary positive result for the corresponding drug of that specific test zone.

Invalid

If a colored band is not visible in each of the Control Region (C) or a colored band is only visible in the Test Region (T), the test is invalid. Another test should be run to re-evaluate the specimen. If the new test still provides an invalid result, please contact the distributor from whom you purchased the product. When calling, be sure to provide the lot number of the test.



Note: There is no meaning attributed to line color intensity or width.

The preliminary positive test result does not always mean that a person took illegal drugs. The negative test result does not always mean that a person did not take illegal drugs. There could be a number of factors that affect the reliability of drug tests. Certain drugs of abuse tests are more accurate than others.

What Is the False Positive Test?

The definition of the false positive test would be the instance where a substance is identified incorrectly by UA Multi-Drug Urine Test Cup. The most common causes of the false positive test are cross reactants. Certain foods and medicines, diet plan drugs and nutritional supplements may cause the false positive test result.

What Is the False Negative Test?

The definition of the false negative test is that the initial drug is present but isn't detected by UA Multi-Drug Urine Test Cup. If the specimen is diluted or adulterated, it may cause the false negative result.

If suspect someone is taking drugs but get the negative test results, please test again at another time, or test for different drugs.

TEST LIMITATIONS

1. This test kit has been developed for testing urine samples only. No other fluids have been evaluated. DO NOT use it to test anything other than urine.
2. It is possible that technical or procedural errors, as well as other interfering substances in the urine specimen may cause false results.
3. This test is a qualitative screening assay. It is not designed to determine the quantitative concentration of drugs or the level of intoxication.

QUESTIONS AND ANSWERS

1. **What does the UA Multi-Drug Urine Test Cup do?**
These tests detect if one or more prescription or illegal drugs such as Amphetamine, Secobarbital, Buprenorphine, Oxazepam, Cocaine, Methyleneiodoxymethamphetamine, Methamphetamine, Morphine, Methadone, Oxycodone, Phencyclidine, Nortriptyline and Cannabinoids are present in urine.

The testing is done in two steps. First, test urine with UA Multi-Drug Urine Test Cup. Second, if any drug test result is preliminary positive, please send the cup with urine to the drug testing laboratory for confirmatory result.

2. **What is "cut-off level"?**

The cut-off level is the specified concentration of a drug in a urine sample. If the concentration of a drug in urine is above the cutoff concentration, this drug test result will be preliminary positive. If the concentration of a drug in urine is below the cutoff concentration, this drug test result will be negative.

3. **What are drugs of abuse?**

Drugs of abuse are illegal or prescription drugs (for example, Oxycodone or Valium) that are taken for a non-medical purpose, including taking the medication longer than doctor prescribed or for a purpose other than what the doctor prescribed.

4. **What are the Common Street Names for the Drugs to be detected?**

Drug	Common Street Names
Amphetamine (AMP)	Speed, Jelly Beans or Super Jellies, Hearts, Uppers, Pick me ups or Wake me ups, Wake ups, Get ups, Boot ups, Sparkles
Secobarbital (BAR)	Amytal, Downers, Nembutal, Phenobarbital, Reds, Red Birds, Red devils, Seconal, Tuninal, Yellowjackets
Buprenorphine (BUP)	Bupe, Subbies, Temmies
Oxazepam (BZO)	Benzos, Downers, Nerve Pills, Tranks
Cocaine (COC)	Blow, C, candy, coke, do a line, freeze, girl, happy dust, Mama coca, mojo, monster, nose, pimp, shot, smoking gun, snow, sugar, sweet stuff, and white powder.
Methyleneiodoxymethamphetamine (MDMA)	Ecstasy, E, X, XTC, Adam, Clarity, Lover's Speed
Methamphetamine (MET/mAMP)	Speed, Ice, Chalk, Meth, Crystal, Crank, Fire, Glass
Methadone (MTD)	mixture, meth, linctus, green
Morphine (MOP/OPI300)	Aunt Hazel, big H, black pearl, brown sugar, capital H, charley, china white, dope, good horse, H, hard stuff, hero, heroina, little boy, mud, perfect high, smack, stuff and tar.
Oxycodone (OXY)	OC, Oycotton, OX, and Kicker
Phencyclidine (PCP)	Angel dust, belladonna, black whack, CJ, cliffhanger, crystal joint, Detroit pink, elephant tranquilizer, hog, magic, Peter Pan, sheets, soma, TAC, trunk, white horizon and zoom.
Nortriptyline (TCA)	Blue angels, Blue birds, Vivacil, Anafranil, Janimine, Tofranil
Cannabinoids (THC)	420, Aunt Mary, baby, bobby, boom, chira, chronic, ditch, ganja, grass, greens, hash, herb, Mary Jane, nigra, Pot, reeper, rip, root, skunk, stack, torch, weed and zambi.

5. **How accurate is the test?**

The tests are sensitive to drugs and accurate. These tests, however, are not as accurate as lab tests. In some cases, certain foods and drugs may cause false positives as well as false negatives for those who use drug testing kits.

6. **If the test results are negative, can the conclusion be that the person is free of drugs?**

This means that if the sample was collected properly and if the test was performed according to direction, then none of the drug screened were present in the urine.

7. **Does a preliminary positive screen test mean that drugs of abuse have been found?**

This means that the test has reacted with something in the urine and the urine must be sent to the lab for a more accurate test.

8. **What should I do, if the lab test confirms a positive result?**

If you have received a confirmed positive result, please consult with counselor for a proper course of action. It is important that you remain calm and do not react in a negative way to the situation. If you do not believe the test result, please consult with your physician. They will have your background medical history and be able to provide you with detailed information on both the test and the meaning of the result.

9. **What is the principle of UA Multi-Drug Urine Test Cup?**

UA Multi-Drug Urine Test Cup is a competitive immunoassay that is used to screen for the presence of drugs of abuse in urine. When the test is activated, the urine is absorbed into the device by capillary action. When flowing across the pre-coated membrane, it will be mixed with the respective drug antibody conjugates. If concentrations of drugs are below corresponding detected drugs' cutoff, respective drug antibody conjugates bind to the respective drug-protein conjugates immobilized in the Test Region (T) of the device. This produces the colored band in Test Region (T) that indicates the negative result. On the contrary, if concentrations of drugs are at or above corresponding detected drugs' cutoff, the free drugs in urine bind to the respective drug antibody conjugates. It prevents the respective drug antibody conjugates from binding to the respective drug-protein conjugates immobilized in the Test Region (T) of the device. Therefore, there is no colored band in the Test Region (T) that indicates the preliminary positive result. To serve as the procedure control, if the test has been performed properly, a colored band will appear at the Control Region (C).

QUALITY CONTROL

Users should follow the appropriate federal, state, and local guidelines concerning the frequency of assaying external quality control materials. Even though there is an internal procedural control line in the test device in the Control Region (C), the use of external controls is strongly recommended as good laboratory testing practice to confirm the test procedure and to verify proper test performance. Positive and negative controls should give the expected results. When testing the positive and negative controls, the same assay procedure should be adopted. External Control (positive and negative) should be run with each new lot of test received, each new shipment, each new operator and monthly to determine that tests are working properly.

PERFORMANCE CHARACTERISTICS

Accuracy

1040 (eighty for each drug) clinical urine specimens were analyzed by GC-MS or LC/MS-MS and by each corresponding UA Multi-Drug Urine Test Cup. Each UA Multi-Drug Urine Test Cup was read by three viewers. Specimens were divided by concentration into five categories: Drug Free, Less than Half the Cutoff, Near Cutoff Negative, Near Cutoff Positive and High Positive. Results were as follows:

Drug Test	UA Cup Result	Drug Free	Less than Half the Cutoff	Near Cutoff Negative (Between 50% below the cutoff and the cutoff)	Near Cutoff Positive (Between the cutoff and 50% above the cutoff)	High Positive (Greater than 50% above the cutoff)	% Agreement with GC/MS or LC/MS
AMP 500	Viewer A	+ 0 0	0	2	30	10	100%
		- 10 17	11	0	0	95%	
		+ 0 0	1	30	10	100%	
		- 10 17	12	0	0	97.5%	
	Viewer B	+ 0 0	0	2	30	10	100%
		- 10 17	11	0	0	95%	
		+ 0 0	1	30	10	100%	
		- 10 17	11	0	0	95%	
	BAR 300	Viewer A	+ 0 0	1	28	11	97.5%
			- 10 19	10	1	0	97.5%
			+ 0 0	1	28	11	97.5%
			- 10 19	10	1	0	97.5%
Viewer B		+ 0 0	1	28	11	97.5%	
		- 10 19	10	1	0	97.5%	
		+ 0 0	1	28	11	97.5%	
		- 10 19	10	1	0	97.5%	
BUP 10		Viewer A	+ 0 0	3	29	10	97.5%
			- 10 18	9	1	0	92.5%
			+ 0 0	3	29	10	97.5%
			- 10 18	9	1	0	92.5%
	Viewer B	+ 0 0	3	29	10	97.5%	
		- 10 18	9	1	0	92.5%	
		+ 0 0	3	29	10	97.5%	
		- 10 18	9	1	0	92.5%	
	BZO 300	Viewer A	+ 0 0	2	29	10	97.5%
			- 10 15	13	1	0	95%
			+ 0 0	2	29	10	97.5%
			- 10 15	13	1	0	95%
Viewer B		+ 0 0	2	29	10	97.5%	
		- 10 15	13	1	0	95%	
		+ 0 0	2	29	10	97.5%	
		- 10 15	13	1	0	95%	
COC 150		Viewer A	+ 0 0	2	31	9	100%
			- 10 18	10	0	0	95%
			+ 0 0	1	31	9	100%
			- 10 18	11	0	0	97.5%
	Viewer B	+ 0 0	2	31	9	100%	
		- 10 18	10	0	0	95%	
		+ 0 0	2	31	9	100%	
		- 10 18	10	0	0	95%	
	MDMA 500	Viewer A	+ 0 0	2	30	10	100%
			- 10 18	10	0	0	95%
			+ 0 0	2	30	10	100%
			- 10 18	10	0	0	95%
Viewer B		+ 0 0	2	30	10	100%	
		- 10 18	10	0	0	95%	
		+ 0 0	2	30	10	100%	
		- 10 18	10	0	0	95%	
Viewer C		+ 0 0	2	20	20	100%	
		- 10 15	13	0	0	95%	
		+ 0 0	2	20	20	100%	
		- 10 15	13	0	0	95%	
MET 500	Viewer A	+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
	Viewer B	+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
	Viewer C	+ 0 0	0	29	10	97.5%	
		- 10 18	12	1	0	100%	
		+ 0 0	0	29	10	97.5%	
		- 10 18	12	1	0	100%	
MOP 300	Viewer A	+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
	Viewer B	+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	27	12	97.5%	
		- 10 18	10	1	0	95%	
	Viewer C	+ 0 0	0	29	10	97.5%	
		- 10 18	12	1	0	100%	
		+ 0 0	0	29	10	97.5%	
		- 10 18	12	1	0	100%	
PCP 25	Viewer A	+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
		+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
	Viewer B	+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
		+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
	Viewer C	+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
		+ 0 0	1	28	10	95%	
		- 10 20	9	2	0	97.5%	
TCA 1000	Viewer A	+ 0 0	2	29	10	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	29	10	97.5%	
		- 10 18	10	1	0	95%	
	Viewer B	+ 0 0	2	29	10	97.5%	
		- 10 18	10	1	0	95%	
		+ 0 0	2	29	10	97.5%	
		- 10 18	10	1	0	95%	
	Viewer C	+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
		+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
THC 50	Viewer A	+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
		+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
	Viewer B	+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
		+ 0 0	3	28	10	95%	
		- 10 19	8	2	0	92.5%	
	Viewer C	+ 0 0	2	29	10	97.5%	
		- 10 19	8	2	0	92.5%	
		+ 0 0	2	29	10	97.5%	
		- 10 19	8	2	0	92.5%	

BAR 300	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	8/42	8/42	8/42
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
	BUP 10	0	50	50/0	50/0
2.5	50	50/0	50/0	50/0	
5	50	50/0	50/0	50/0	
7.5	50	50/0	50/0	50/0	
10	50	9/41	9/41	10/40	
12.5	50	0/50	0/50	0/50	
15	50	0/50	0/50	0/50	
17.5	50	0/50	0/50	0/50	
20	50	0/50	0/50	0/50	
BZO 300	0	50	50/0	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	7/43	7/43	8/42
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
	COC 150	0	50	50/0	50/0
37.5	50	50/0	50/0	50/0	
75	50	50/0	50/0	50/0	
112.5	50	50/0	50/0	50/0	
150	50	10/40	11/39	10/40	
187.5	50	0/50	0/50	0/50	
225	50	0/50	0/50	0/50	
262.5	50	0/50	0/50	0/50	
300	50	0/50	0/50	0/50	
MDMA 500	0	50	50/0	50/0	50/0
	125	50	50/0	50/0	50/0
	250	50	50/0	50/0	50/0
	375	50	50/0	50/0	50/0
	500	50	11/39	10/40	10/40
	625	50	0/50	0/50	0/50
	750	50	0/50	0/50	0/50
	875	50	0/50	0/50	0/50
	1000	50	0/50	0/50	0/50
	MET 500	0	50	50/0	50/0
125		50	50/0	50/0	50/0
250		50	50/0	50/0	50/0
375		50	50/0	50/0	50/0
500		50	10/40	9/41	9/41
625		50	0/50	0/50	0/50
750		50	0/50	0/50	0/50
875		50	0/50	0/50	0/50
1000		50	0/50	0/50	0/50
MOP 300		0	50	50/0	50/0
	75	50	50/0	50/0	50/0
	150	50	50/0	50/0	50/0
	225	50	50/0	50/0	50/0
	300	50	10/40	11/39	10/40
	375	50	0/50	0/50	0/50
	450	50	0/50	0/50	0/50
	525	50	0/50	0/50	0/50
	600	50	0/50	0/50	0/50
	MTD 300	0	50	50/0	50/0

	75	50	50/0	50/0	50/0	
	150	50	50/0	50/0	50/0	
	225	50	50/0	50/0	50/0	
	300	50	9/41	9/41	8/42	
	375	50	0/50	0/50	0/50	
	450	50	0/50	0/50	0/50	
	525	50	0/50	0/50	0/50	
	600	50	0/50	0/50	0/50	
	OXY 100	0	50	50/0	50/0	50/0
	25	50	50/0	50/0	50/0	
50	50	50/0	50/0	50/0		
75	50	50/0	50/0	50/0		
100	50	9/41	9/41	9/41		
125	50	0/50	0/50	0/50		
150	50	0/50	0/50	0/50		
175	50	0/50	0/50	0/50		
200	50	0/50	0/50	0/50		
PCP 25	0	50	50/0	50/0	50/0	
	6.25	50	50/0	50/0	50/0	
	12.5	50	50/0	50/0	50/0	
	18.75	50	50/0	50/0	50/0	
	25	50	7/43	6/44	7/43	
	31.25	50	0/50	0/50	0/50	
	37.5	50	0/50	0/50	0/50	
	43.75	50	0/50	0/50	0/50	
	50	50	0/50	0/50	0/50	
	0	50	50/0	50/0	50/0	
TCA 1000	250	50	50/0	50/0	50/0	
	500	50	50/0	50/0	50/0	
	750	50	50/0	50/0	50/0	
	1000	50	10/40	10/40	10/40	
	1250	50	0/50	0/50	0/50	
	1500	50	0/50	0/50	0/50	
	1750	50	0/50	0/50	0/50	
	2000	50	0/50	0/50	0/50	
	0	50	50/0	50/0	50/0	
	THC 50	12.5	50	50/0	50/0	50/0
25		50	50/0	50/0	50/0	
37.5		50	50/0	50/0	50/0	
50		50	10/40	10/40	11/39	
62.5		50	0/50	0/50	0/50	
75		50	0/50	0/50	0/50	
87.5		50	0/50	0/50	0/50	
100		50	0/50	0/50	0/50	

Specificity and Cross Reactivity

To test the specificity, the test device was used to test various drugs, drug metabolites and other components of the same class that are likely to be present in urine. All the components were added to drug-free normal human urine. The following structurally related compounds produced positive results with the test when tested at levels equal to or greater than the concentrations listed below.

Substance	Conc. (ng/mL)	Substance	Conc. (ng/mL)
AMP 500			
d-Amphetamine	500	l-Amphetamine	25,000
d,l-Amphetamine	1,500	(+/-) 3,4-Methylenedioxyamphetamine (MDA)	2,500
Phentermine	1,500	Hydroxyamphetamine	8,000
d-methamphetamine	>100,000	l-methamphetamine	>100,000
(+/-) 3,4-Methylenedioxyethylamphetamine (MDE)	>100,000	Methylenedioxyamphetamine (MDMA)	>100,000

Ephedrine	>100,000	β -Phenylethylamine	100,000
Tyramine	100,000	p-Hydroxynorephedrine	100,000
Phenylpropanolamine	>100,000	(\pm) Phenylpropanolamine	>100,000
d/l-Norephedrine	100,000	Benzphetamine	>100,000
l-Ephedrine	>100,000	l-Epinephrine	>100,000
d/l-Epinephrine	>100,000		
BAR 300			
Secobarbital	300	Butathal	100
Amobarbital	10,000	Butalbital	2,500
Alphenol	150	Cyclopentobarbital	600
Aprobarbital	200	Pentobarbital	2,500
Butobarbital	75	Phenobarbital	10,000
BUP 10			
Buprenorphine	10	Norbuprenorphine	20
Buprenorphine-3-D-Glucuronide	15	Norbuprenorphine-3-D-Glucuronide	200
Morphine	>100,000	Oxymorphone	>100,000
Hydromorphone	>100,000		
BZO 300			
Oxazepam	300	Diazepam	200
Alprazolam	200	Estazolam	1,000
α -Hydroxalprazolam	1,500	Flunitrazepam	2,500
Bromazepam	500	D,L-Lorazepam	1,500
Chlordiazepoxide	1,500	Midazolam	12,500
Clobazam	100	Nitrazepam	4,000
Clonazepam	800	Norchlordiazepoxide	200
Clorazepate dipotassium	200	Nordiazepam	500
Delorazepam	1,500	Temazepam	250
Desalkylfurazepam	400	Triazolam	1,200
Demoxepam	2,000	Flurazepam	500
COC 150			
Benzoyllecgonine	150	Ecgonine	16,000
Cocaine	375	Ecgonine methyl ester	>100,000
Cocaethylene	6,250	Norcocaine	>100,000
MDMA 500			
3,4-Methylenedioxyamphetamine (MDMA)	500	3,4-Methylenedioxyethylamphetamine (MDEA)	300
3,4-Methylenedioxyamphetamine (MDA)	3,000	d-Methamphetamine	>100,000
l-methamphetamine	50,000	l-amphetamine	>100,000
d-amphetamine	>100,000		
MET 500			
d-methamphetamine	500	(+/-) 3,4-Methylenedioxyamphetamine (MDMA)	2,000
p-Hydroxymethamphetamine	15,000	(-)-Methamphetamine	12,500
l-methamphetamine	10,000	d-Amphetamine	25,000
l-Amphetamine	37,500	Chloroquine	10,000
(+/-)-Ephedrine	25,000	d/l-Methamphetamine	500
L-Methamphetamine	10,000	(+/-) 3,4-Methylenedioxyethylamphetamine (MDEA)	500
(+/-) 3,4-Methylenedioxyamphetamine (MDA)	500	β -Phenylethylamine	25,000
Trimethobenzamide	5,000	d/l-Amphetamine	75,000
p-Hydroxymethamphetamine	15,000	Mephentermine	25,000
(1R,2S)-(-)-Ephedrine	50,000	l-Phenylephrine	100,000
MOP 300			
Morphine	300	Morphine-3- β -d-glucuronide	1,000
Codeine	300	Norcodeine	6,250

Ethyl Morphine	100	Normorphine	300
Heroin	300	Oxycodone	10,000
Hydrocodone	5,000	Oxymorphone	10,000
Hydromorphone	1,000	Procaine	150,000
6-Monoacetylmorphine (6-MAM)	150	Thebaine	3,000
Levorphanol	10,000		
MTD 300			
Methadone	300	Doxylamine	50,000
EMDP	>100,000	EDDP	>100,000
LAAM	>100,000	Alpha Methadol	>100,000
OXY 100			
Oxycodone	100	Codeine	100,000
Dihydrocodeine	20,000	Ethyl Morphine	>100,000
Hydrocodone	10,000	Hydromorphone	32,000
Oxymorphone	1,000	Thebaine	>100,000
Acetylmorphine	>100,000	Morphine	>100,000
Buprenorphine	>100,000		
PCP 25			
Phencyclidine	25	4-Hydroxyphencyclidine	12,500
TCA 1000			
Nortriptyline	1,000	Promazine	1,500
Amitriptyline	1,500	Maprotiline	2,000
Clomipramine	12,500	Nordoxepin	1,000
Desipramine	200	Promethazine	25,000
Doxepin	2,000	Trimipramine	3,000
Imipramine	400	Cyclobenzaprine	800
Norclomipramine	12,500		
THC 50			
11-nor- Δ 9-THC-9-COOH	50	Δ 9-Tetrahydrocannabinol	5,000
11-nor- Δ 8-THC-9-COOH	30	Cannabinol	20,000
11-hydroxy- Δ 9-Tetrahydrocannabinol	5,000	Cannabidiol	100,000
Δ 8-Tetrahydrocannabinol	1,300	11-nor- Δ 9-THC-carboxy-glucuronide	100

Effect of Urinary Specific Gravity

The results demonstrate that the urinary specific gravity range of 1.000~1.035 does not affect the test results.

Effect of Urinary pH

The results demonstrate that the range of urinary pH from 4 to 9 does not interfere with the performance of test.

Interfering Substances

The following compounds were added to drug-free urine, urine with drug concentration 25% below the cutoff, and urine with drug concentration 25% above the cutoff for the corresponding UA Multi-Drug Urine Test Cup. All potential interferents were added at a concentration of 100 μ g/mL. None of the urine samples showed any deviation from the expected results.

(-) Cotinine	Ecgonine Methyl Ester	Nimodipine
3-Hydroxytyramine	Effexor	Norethindrone
Acetaminophen	Enalapril Maleate	O-Hydroxyhippuric Acid
Acetophenetidin	Epinephrine Hydrochloride	Olanzapine
Acetylsalicylic Acid	Erythromycin	Omeprazole
Acyclovir	Esomeprazole Magnesium	Ondansetran
Afrin	Ethanol	Oxalic Acid
Albumin	Fenofibrate	Oxolinic Acid
Aminophylline	Fenoprofen	Oxymetazoline
Aminopyrine	Fentanyl Citrate	Paliperidone
Amiodarone Hydrochloride	Fluoxetine Hydrochloride	Pantoprazole

Amlodipine Mesylate	Fluvoxamine	Papaverine
Amoxicillin	Furosemide	Paroxetine Hydrochloride
Ampicillin	Gabapentin	Penfluridol
Apomorphine	Geniticis Acid	Penicillin-G
Aripiprazole	Glubendamide	Penicillin V Potassium
Aspartame	Glilazide	Phenelzine
Atomoxetine	Glipizide	Pioglitazone Hydrochloride
Atorvastatin Calcium	Glucose	Piracetam
Atropine	Haloperidol	Pravastatin Sodium
Benzoic Acid	Hemoglobin	Prednisone
Benzoic Acid	Ibuprofen	Promethazine
Bilirubin	Isosorbide Dinitrate	Propylthiouracil
Bupropion	Isosuprine	Quetiapine Fumarate
Captopril	Ketamine	Quinine
Carbamazepine	Ketoconazole	Ranitidine
Cefradine	Ketoprofen	Rifampicin
Cephalexin	Kratom	Risperidone
Chloral Hydrate	Labeltalol	Salicylic Acid
Chloramphenicol	Lamotrigine	Serotonin
Chloroquine		